



Unit 4C | Woodhall Business Park | Sudbury | Suffolk | CO10 1WH | United Kingdom
T: 01787 882038 | F: 01787 882039 | E: info@modusgauges.com | W: modusgauges.com

STANDARD CALIBRATION CERTIFICATE NO: 48973 -1

Instrument Serial No: 48973 -1

Instrument Description
100MM DIAMETER METAL CASE BRASS INTERNALS BOTTOM ENTRY 1/2 BSPT THREAD 0-1 BAR/PSI
Part No: B401A240DT

We certify that the above instrument has been tested to BS EN 837-1: 1998.
Accuracy tested to and within: \pm 1.60% FSD Limits

Instrument tested with: AIR
Instrument tested in: VERTICAL position
Laboratory Temperature 20°C \pm 5°C

Calibration Equipment Used: Digital Gauge: 0-2.5 Bar
Serial No: 110811304D
Accuracy: 0.05%

Calibration equipment above is within its period of calibration and traceable to NPL/National Standards

Signed: 

Date: 06/02/2020