



Unit 4C | Woodhall Business Park | Sudbury | Suffolk | CO10 1WH | United Kingdom
T: 01787 882038 | F: 01787 882039 | E: info@modusgauges.com | W: modusgauges.com

STANDARD CALIBRATION CERTIFICATE NO: 49343 -1

Instrument Serial No: 49343 -1

Instrument Description
63MM DIAMETER STAINLESS STEEL CASE BRASS INTERNALS BOTTOM ENTRY 1/4 BSP THREAD 0-60 PSI/BAR GLYCERINE FILLED
Part No: B222A244K

We certify that the above instrument has been tested to BS EN 837-1: 1998.
Accuracy tested to and within: \pm 1.60% FSD Limits

Instrument tested with: AIR
Instrument tested in: VERTICAL position
Laboratory Temperature 20°C \pm 5°C

Calibration Equipment Used: Digital Gauge: 0-10 Bar
Serial No: 110911392D
Accuracy: 0.05%

Calibration equipment above is within its period of calibration and traceable to NPL/National Standards

Signed:

Date: 27/07/2020



Unit 4C | Woodhall Business Park | Sudbury | Suffolk | CO10 1WH | United Kingdom
T: 01787 882038 | F: 01787 882039 | E: info@modusgauges.com | W: modusgauges.com

STANDARD CALIBRATION CERTIFICATE NO: 49343 -2

Instrument Serial No: 49343 -2

Instrument Description
63MM DIAMETER STAINLESS STEEL CASE BRASS INTERNALS BOTTOM ENTRY 1/4 BSP THREAD 0-60 PSI/BAR GLYCERINE FILLED
Part No: B222A244K

We certify that the above instrument has been tested to BS EN 837-1: 1998.
Accuracy tested to and within: \pm 1.60% FSD Limits

Instrument tested with: AIR
Instrument tested in: VERTICAL position
Laboratory Temperature $20^{\circ}\text{C} \pm 5^{\circ}\text{C}$

Calibration Equipment Used: Digital Gauge: 0-10 Bar
Serial No: 110911392D
Accuracy: 0.05%

Calibration equipment above is within its period of calibration and traceable to NPL/National Standards

Signed:

Date: 27/07/2020