GAUGES LIMITED

Unit 4C | Woodhall Business Park | Sudbury | Suffolk | CO10 1WH | United Kingdom
T: 01787 882038 | F: 01787 882039 | E: info@modusgauges.com | W: modusgauges.com

STANDARD CALIBRATION CERTIFICATE NO: 49433 -1

Instrument Serial No: 49433 -1

Instrument Description
100MM DIAMETER
STAINLESS STEEL CASE
BRASS INTERNALS
BOTTOM ENTRY
1/2 BSP THREAD
0-15 PSI/BAR
GLYCERINE FILLED
Part No: B422A240M

We certify that the above instrument has been tested to BS EN 837-1: 1998.
Accuracy tested to and within: ± 1.60% FSD Limits

Instrument tested with: AIR
Instrument tested in: VERTICAL position
Laboratory Temperature 20°C ± 5°C

Calibration Equipment Used:

Digital Gauge:	0-2.5 Bar
Serial No:	110811304D
Accuracy:	0.05%

Calibration equipment above is within its period of calibration and traceable to NPL/National Standards

Signed: K Greenwood Date: 24/08/2020