

REPAIR No: 20-004111

CERT No: 2012009

DATE: 3rd December 2020

CALIBRATION CERTIFICATE

INSTRUMENT TYPE: 12.5 tonnef LLX

SERIAL NUMBER: F00064

CALIBRATED BY D Rodman

ON: 3rd December 2020

CALIBRATED AT S & S ELECTRONICS

USING: Tractel MasterLoad Link No.2
Serial No. 83 AA3

Temperature 20.2 °C

MASTER LOAD LINK No. 2 CALIBRATED ON 23rd June 2020 EXPIRES ON 24th JUNE 2021 CERTIFICATE NO: 2006036

BY R & H TESTING SERVICES LTD (UKAS APPROVAL 0252)

TRUE LOAD	ACTUAL READING	ERROR
kgf	kgf	- / +

LINEARITY CHECK


	0	0	
Calibrated to 0.2 % accuracy of full scale i.e. plus or minus 25 kgs at any load between 0-12.5 TONNES.	1250.0	1235.0	-15.0
Traceable records on our Force Machine in accordance BS EN ISO 7500-1:2018	2500.0	2485.0	-15.0
	5000.0	4975.0	-25.0
	7500.0	7480.0	-20.0
	10000.0	9980.0	-20.0
	12500.0	12490.0	-10.0

Hand Held indicator reads same as above.

ISSUED TO: Generator Installations

For and on behalf of
TRACTEL (UK) LTD

Form 41 Issue 5 Oct 2020



D Rodman



Tractel (UK) Ltd | Old Lane | Halfway | Sheffield S20 3GA
Registered in UK No. 533669 | VAT Reg No. GB 173 1249 76
E-mail sales.uk@tractel.com | Phone +44 114 248 2266



Connect with us:  tractel

 @tractel.official

 tractel-uk-limited

 @tractelofficial

 tractel