



Unit 4C | Woodhall Business Park | Sudbury | Suffolk | CO10 1WH | United Kingdom
T: 01787 882038 | F: 01787 882039 | E: info@modusgauges.com | W: modusgauges.com

STANDARD CALIBRATION CERTIFICATE NO: 49899 -2

Instrument Serial No: 49899 -2

Instrument Description
63MM DIAMETER STAINLESS STEEL CASE BRASS INTERNALS BOTTOM ENTRY 1/4 BSP THREAD 0-200 PSI/BAR GLYCERINE FILLED
Part No: B222A250K

We certify that the above instrument has been tested to BS EN 837-1: 1998.
Accuracy tested to and within: ± 1.60% FSD Limits

Instrument tested with: WATER
Instrument tested in: VERTICAL position
Laboratory Temperature 20°C ± 5°C

Calibration Equipment Used:	Digital Gauge:	0-16 Bar
	Serial No:	10458247
	Accuracy:	0.125%

Calibration equipment above is within its period of calibration and traceable to NPL/National Standards

Signed: 

Date: 10/02/2021